



खडक नगरपालिका

नगर कार्यपालिकाको कार्यालय

कल्याणपुर, सप्तरी

मधेश प्रदेश, नेपाल



पत्र संख्या : २०८२/०८३

च.नं. १४६४

मिति: २०८३/०९/२८

विषय : इस्टिमेट प्रयोजनका लागि दररेट माँग सम्बन्धमा ।

(प्रथम पटक प्रकाशित मिति: २०८३/०९/२८)

प्रस्तुत विषयमा आ.व. २०८२/८३ मा खडक नगरकार्यपालिकाको आधारभूत अस्पताल कल्याणपुरलाई आवश्यक पर्ने सरसामानहरूको दररेट आवश्यक परेको हुँदा सम्बन्धित विषयमा कारोबार गर्ने इच्छुक, दर्ता भएका फर्म/कम्पनी/ आपूर्तिकर्ताहरूले यो सूचना प्रकाशित भएको मितिले ७ (सात) दिन भित्र तपसिल बमोजिमका कागजातहरू सहित दररेट पेश गर्नु हुन यो सूचना प्रकाशित गरिएको छ । तोकिएको समयभित्र प्राप्त नभएका दररेटहरू स्वीकार गरिने छैनन् । प्राप्त दररेटहरू इस्टिमेट प्रयोजनका लागि मात्र प्रयोग गरिने व्यहोरा समेत आदेशानुसार जानकारी गराइन्छ । साथै सरसामानहरूको सूची यस खडक नगरपालिकाको website: khadakmun.gov.np मा डाउनलोड गर्न सकिनेछ ।

तपसिल:

१. फर्म/संस्था दर्ता प्रमाणपत्रको प्रतिलिपी
२. स्थायी लेखा नं./मु.अ.कर दर्ता प्रमाणपत्रको प्रतिलिपी
३. आ.व. २०८१/०८२ को करचुक्ता प्रमाणपत्रको प्रतिलिपी

(हरेन्द्र कुमार श्रेष्ठ)

प्रमुख प्रशासकीय अधिकृत
हरेन्द्र कुमार श्रेष्ठ
प्रमुख प्रशासकीय अधिकृत

S.N	Name Of Items/Particulars		Unit	Quantity	Rate (Excluding Vat)	Remarks
1	Protein Analyzer With Chemical	As Per Attached Specifications	PCS	1		
	ASO		IU/mL	1		
	CRP		mg/L	1		
	RF		IU/ML	1		
	HbA1c		%	1		
	Microalbumin		mg/L	1		
	Ferritin		ng/mL	1		
	Cystatin C		mg/L	1		
	IgM		Mg/dL	1		
	hsCRP		mg/L	1		
	IgE		IU/ML	1		
	IgG		Mg/dL	1		
	IgA		mg/dL	1		
	Lp(a)		Mg/dL	1		
	Apo A1		Mg/dL	1		
	Apo B		Mg/dL	1		
	C3		Mg/dL	1		
	C4		Mg/dL	1		
	Haptoglobin		Mg/dL	1		
	D-Dimer		Mg/L	1		
	Ceruloplasmin		mg/dl	1		
	H.Pylori		U/mL	1		
	Vitamin D		ng/mL	1		
	PCT		ng/mL	1		
	Wash Solution		ml	1		
	Wash Solution 100x		ml	1		
	HbA1c Control 1 & 2		ml	1		
	Protein Control (Bi level)		ml	1		
	Microalbumin Control (Bi Level)		ml	1		
	D-Dimer Control (Bi level)		ml	1		
	Cystatin Control (Bi level)	ml	1			
	HS CRP Control (Bi level)	ml	1			
	Vitamin D Control (Bi level)	ml	1			
2	Physiotherapy (Thermal Massager)		PCS	1		

15/10/23

Protein Analyser

S.N.	Purchaser's Specifications	Bidder's Offer
	Proteins Analyzer	
	Manufacturer	
	Brand	
	Type / Model	
	Country of Origin	
1	Description of Function	
1.1	Quantitative determination of specific proteins in blood, serum, urine samples.	
2	Operational Requirements	
2.1	Protein Analyzer complete with printer is required.	
3	System Configuration	
3.1	Fixed time Nephelometry.	
4	Technical Specifications	
4.1	Should support cartridge-based or equivalent system for ease of use	
4.2	Colour screen display.	
4.3	System should be capable of performing the following parameters: hs-CRP, CRP, ASO, RF, IgG, IgM, IgA, C3, C4, Cystatin-C and microalbumin (mALB) or equivalent.	
4.4	Sample Volume: 5ul – 20ul	
4.5	Sample type: Serum, plasma, urine	
4.6	System should have automatic cleaning and low maintenance requirements	
4.7	Reagents should be identifiable with minimal manual intervention	
4.8	System should have minimal carryover ($\leq 0.1\%$ or as per international standards)	
4.9	System should ensure minimal carryover ($\leq 0.1\%$) using disposable tips or effective automated washing system	
6	Operating Environment	
6.1	The system offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Power Supply, Climate, Temperature, Humidity, etc.	
6.2	Power supply: 220 – 240VAC, 50Hz fitted with appropriate plug. The power cable must be at least 3m in length.	
7	Standards and Safety Requirements	
7.1	CE (93/42 EEC Directives) or UL or TUV or USFDA approved product certificate	
7.2	Must be compliant to ISO 13485	
7.3	Shall meet IEC 61010-1 safety requirements for electrical equipment for measurement, control, and laboratory use.	
8	User Training	
8.1	Must provide user training (including how to use and maintain the equipment).	
9	Warranty	
9.1	Comprehensive warranty for 2 years after acceptance.	
10	Maintenance Service During Warranty Period	

Handwritten signature




S.N.	Purchaser's Specifications	Bidder's Offer
10.1	During the warranty period supplier must ensure corrective/breakdown maintenance whenever required.	
11	Installation and Commissioning	
11.1	Supplier must accomplish proper installation and commissioning of the equipment on site.	
12	Documentation	
12.1	User (Operating) manual in English	
12.2	Service (Technical / Maintenance) manual in English	
12.3	List of important spare parts and accessories with their part numbers and costing.	
12.4	Certificate of calibration and inspection from factory.	




Thermal Massager Therapy

S.N.	Purchaser's Specifications	Bidder's Offer
	Thermal Massager Therapy	
	Manufacturer	
	Brand	
	Type/Model	
	Country of Origin	
1	Description of Function	
1.1	A thermal massager combines massage therapy & heat therapy to help relax muscles, improve circulation, and reduce stiffness or soreness. They are commonly used for the neck, shoulders, back, legs, feet, and spine.	
2	Operational Requirements	
2.1	A thermal massager used for the neck, shoulders, back, legs, feet, and spine	
3	System Configuration	
3.1	A thermal massager with complete accessories	
4	Technical Specifications	
4.1	Internal Projector: It shall have optimum size and angle that ensures acupressure is applied is applied uniformly throughout the body	
4.2	External Projector: It shall be used on any desired area of the body	
4.3	It shall have remote function	
4.4	It shall have adjustable length	
4.5	It shall have mat in the bottom part that can be folded	
5	Accessories, spares and consumables	
5.1	All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication materials, to be included in the offer. Bidders must specify the quantity of every item included in their offer (including items not specified above).	
6	Operating Environment	
6.1	The system offered shall be designed to store and to operate normally under the conditions of the purchaser's country. The conditions include Power Supply, Climate, Temperature, Humidity, etc.	
6.2	Power supply: 220-240V/ 50 Hz AC Single phase fitted with appropriate plug to meet purchaser's country requirements. The power cable must be minimum 3 metres long.	
7	Standards and Safety Requirements	
7.1	Must submit ISO13485:2007	
7.2	CE or USFDA approved product certificate.	
8	User Training	
8.1	Must provide user training (including how to use and maintain the equipment).	
9	Warranty	
9.1	Comprehensive warranty for 2 years.	
10	Maintenance Service During Warranty Period	
10.1	During the warranty period supplier must ensure corrective/breakdown maintenance whenever required.	
11	Installation and Commissioning	
11.1	Supplier must accomplish proper installation & commissioning of the equipment onsite.	
12	Documentation	
12.1	User (Operating) manual in English.	
12.2	Service (Technical / Maintenance) manual in English.	






 ಹೆಚ್.ಎಸ್.ಎಸ್. ಕೆ.ಎಸ್.ಎಸ್.ಎಸ್.

 (ಇ.ಎಸ್.ಎಸ್.)

 ೨೦/೧೨ ೨೦೨೩

S.N.	Purchaser's Specifications	Bidder's Offer
12.3	List of important spare parts and accessories with their part numbers and costing.	
12.4	Certificate of calibration and inspection from factory.	




 डा. शैलेन्द्र कुमार
 वा. से. ट. (डि.)
 प्र. स्वा. आ. व्या. के. जनकपुर